



DR NAOURI SETTINGS FOR SECRET

Pre-care procedure

- EMLA cream every 10min 2 hours prior to the treatment
- Thoroughly cleanse with an antiseptic wipes or solution

Intensive Face care procedure

- Adjust the needle depth according to the skin irregularities and surface.
- Start at 2.5mm on the thickest or laxest areas (cheeks, neck...)
- Reduce depth when needles hit on a bone contact or on areas with thinner skin (hands, forehead....) by .5 to .5 until you find the maximum needle depth without any bone contact. (not less than 1mm depth).
- Intensity : 60 in general
- RF duration : proportional to depth : 2,5mm depth → 250ms ; 1,5mm depth → 150 ms...
- Possibility to increase RF duration by 50ms (1mm depth → 150ms...) but for expert users only.
- Treatment frequency : 1Hz (adjustable)
- Delay time : 2X the RF duration time. (100 ms → 200 ms ; 250ms → 500 ms...)
- 2 passes minimum with 50% overlap horizontally and vertically
- Minimum number of spots : 500 for entire face / 300 for cheeks

Other indications :

Scars : If possible use the 3.5 mm depth – 5 passes slightly moving the spots to avoid stacking

Naso-labial fold and deep wrinkles : If possible use the 3.5 mm depth – 3 to 5 passes overlapping the wrinkle

Hyperhidrosis : 80 / 3,5 / 350 / 2 passes. 1st pass horizontal / 2nd pass vertical – Minimum 150 spots/axilla

Stretch marks : 60/ 3.5 / 350 for the 1st pass.
60/ 2.5 / 200 for the 2nd pass
60 / 1,5/ 100 for the 3rd pass

Post-care : Soothing cream on the treated areas.