

## 64 SEMI-INSULATED NEEDLING I PROTOCOL GUIDLINE

Updated: 2019-09-24

Provided by Michael Naouri, M.D, France

Main Indication	Treatment Area	Intensity(%)	RF Duration(ms)	Depth(mm)	Delay Time(ms)	Pass
Scar / Rhytids	Forehead	20 - 40	100 - 120	1 - 1.5	300	2
	Nose	20 - 40	100 - 120	1 - 1.5	300	2
	Periorbital	20 - 40	100 - 120	1 - 1.5	300	2
	Cheek	40 - 60	150	1.5	300	2
	Lip	40 - 60	150	1.5	300	2
	Chin	40 - 60	150	1.5	300	2
	Neck	20 - 40	150	1.5	300	2
	Body	40 - 60	150	1.5	300	2

## \*\* MOST IMPORTANT THING TO AVOID COMPLICATIONS: FOLLOW THE CAUTION BELOW AND DON'T MOVE THE TIP TOO FAST!!

- $\sqrt{}$  The tip head must be firmly pressed on the skin before pressing the foot pedal and until the needles are completely retracted.
- $\sqrt{\mbox{ Use single mode}}$  and press on the foot pedal per each shoot.
- $\sqrt{\text{Never be faster than 1sec}}$  and be very careful with interval mode.
- $\sqrt{\,}$  Delay time must be more than twice from the value of RF duration.
- $\sqrt{\text{Depth must be adjusted to insulation of the needles: more than 1mm.}}$
- $\sqrt{}$  If it smells burn, stop treatment or decrease applied parameters.
- $\sqrt{\text{Never}}$  use more than 60% intensity except hyperhidrosis.



<sup>\*\*</sup> Intensive treatment for cheek, lip, chin & body you can use; 40-60% intensity – 250ms RF duration – 2.5mm depth – 500ms delay time. Secret RF tips are single-use only.